



Fulton County, GA

Department of Purchasing & Contract Compliance

November 6, 2015

Re: 15ITB98501A-CJC, Digital X-Ray System & Installation/Tuberculosis (TB) Clinic

Dear Proposers:

Attached is one (1) copy of Addendum 2, hereby made a part of the above referenced ITB.

Except as provided herein, all terms and conditions in the ITB referenced above remain unchanged and in full force and effect.

Sincerely,

Charlie Crockett
Assistant Purchasing Agent

Winner 2000 - 2009 Achievement of Excellence in
Procurement Award • National Purchasing Institute



**15ITB98501A-CJC, Digital X-Ray System & Installation/Tuberculosis (TB) Clinic
Addendum No. 2
Page Two**

This Addendum forms a part of the contract documents and **modifies** the original RFP documents as noted below:

- See Attachment#2 for changes in Section 3 of the SCOPE OF WORK AND TECHNICAL SPECIFICATIONS
- See Attachment#1 for changes in Section 7 of the PRICING FORM

ACKNOWLEDGEMENT OF ADDENDUM NO. 2

The undersigned proposer acknowledges receipt of this addendum by returning one (1) copy of this form with the proposal package to the Department of Purchasing & Contract Compliance, Fulton County Public Safety Building, 130 Peachtree Street, Suite 1168, Atlanta, Georgia 30303 by the ITB due date and time **Thursday, November 12, 2015 11:00 A.M.**

This is to acknowledge receipt of Addendum No. 2, _____ day of _____, 20__.

Legal Name of Bidder

Signature of Authorized Representative

Title

Attachment#1

SECTION 3

SCOPE OF WORK AND TECHNICAL SPECIFICATIONS

The Fulton County Department of Health and Wellness is requesting Bids from qualified vendors to provide a digital imaging integrated system for the Tuberculosis (TB) Clinic. The TB Clinic would like to convert to a digital radiography system which will require a complete overhaul of the room and equipment. The dimensions of the x-ray room is 22'L x 14'W. The current equipment is analog and is not compatible with digital data, therefore it needs to be replaced. We are interested in purchasing a system with software which we will be able to view images instantaneously, and view in the Mitchell and McCormick (M&M)/EHR along with radiology reports.

A representative list of the requirements is attached as Price Sheet. The list will be used for comparison of quotes for the purpose of awarding the contract. Items must be equal to or compatible brands.

The successful vendor shall provide the following:

Digital Imaging System –

1. ***Viztek Straight Arm System Positioner/Swivel Arm or equivalent***
2. ***DR Detector***
3. ***High Frequency Generator / X-Ray Console***
4. ***50KW Generator / 25 KHz***
5. ***X-Ray Tube***
6. ***Mobile X-ray table***
7. ***Collimator***
8. ***Acquisition Workstation***
9. ***Diagnostic Workstations (2)***
10. ***Film Digitizer System to scan existing medical X-Ray films and reports into an electronic format. Film digitizer system to include equipment to store and retrieve images.***
11. ***Software to view images once they are taken.***
12. ***Shielding design and assistance with state registration process, to include initial inspection.***

Pricing Form

Item No.	Item Description	Model No.	Estimated Quantity	Unit of Issue	Unit Price
<p>50KW Generator / 25 KHz</p>	<p><u>High Frequency Generator / X-Ray Console</u></p> <p>50KW Generator / 25 KHz / KV Range 40 – 150KVp in 1 KVp step / mA Range 10 – 640mA in 19 steps, Renard scale/ mAs range -0.1 to 500 mAs in 38 step, Renard scale /Exposure time range -0.001 – 10 seconds Output power – 640 mA @ 100 KVp / 500 mA @ 125 KVp / 400 mA @ 150 KVp</p>	<p>SHF525</p>	<p>1</p>	<p>EA</p>	<p>\$ _____</p>